



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/784,906	02/23/2004	Hugo X. Gonzalez	SPIRTN.017C1	6137
20995	7590	06/05/2006	EXAMINER	
KNOBBE MARTENS OLSON & BEAR LLP 2040 MAIN STREET FOURTEENTH FLOOR IRVINE, CA 92614			FOREMAN, JONATHAN M	
			ART UNIT	PAPER NUMBER
			3736	

DATE MAILED: 06/05/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/784,906

Applicant(s)

GONZALEZ, HUGO X.

Examiner

Jonathan ML Foreman

Art Unit

3736

-- The MAILING DATE of this communication appears on the cover sheet with the correspond nc addr ss --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 3/13/06.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 11-18 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 11-18 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 11 – 13 and 16 – 18 are rejected under 35 U.S.C. 102(b) as being anticipated by US Patent No. 2,115,298 to Brown.

In regards to claims 11 – 13 and 16 - 18, Brown discloses an elongate body (Figures 1) having a proximal end (2x), a distal end (3), and a lumen (2) therethrough, the proximal end being configured to couple with a vacuum source to allow a vacuum to be selectively drawn through the lumen of the elongate body (Col. 2, lines 1 - 5), the distal end configured to contact a body tissue sample prior to resection (Col. 1, lines 49 – 50) and is capable of housing a collection bag. Brown discloses a stop mechanism (20) configured to selectively reduce an inner diameter of the lumen and means for activating (18) the stop mechanism. A recitation with respect to the manner in which an apparatus is intended to be employed does not impose any structural limitation upon the claimed apparatus that differentiates it from a prior art reference disclosing the structural limitations of the claim. *In re Pearson*, 494 F.2d 1399, 181 USPQ 641 (CCPA 1947); *In re Yanush*, 477 F.2d 958, 177 USPQ 705 (CCPA 1973); *In re Finsterwalder*, 436 F.2d 1028, 168 USPQ 530 (CCPA 1971); *In re Casey*, 370 F.2d 576, 152 USPQ 235 (CCPA 1967); *In re Otto*, 312 F.2d 937, 136 USPQ 458 (CCPA 1963); *Ex parte Masham*, 2 USPQ2d 1647 (BbPatApp & Inter 1987). In the present case, the apparatus as claimed intended to house a collection bag. The claim language “configured to house a collection bag” does not require the collection bag to be part of the apparatus, but merely capable of being

Art Unit: 3736

associated with the apparatus. Because there is no structure associated with the apparatus as disclosed by Brown which would prevent a collection bag from being placed within the apparatus, the apparatus is “configured to house a collection bag”.

3. Claims 11, 12 and 14 – 17 are rejected under 35 U.S.C. 102(b) as being anticipated by US Patent No. 5,281,229 to Neward.

In regards to claims 11, 12 and 14 – 17, Neward discloses an elongate body (Figures 1 and 2) having a proximal end (16), a distal end (10), and a lumen therethrough, the proximal end being configured to couple with a vacuum source to allow a vacuum to be selectively drawn through the lumen of the elongate body (Col. 2, lines 33 – 39), the distal end configured to contact a body tissue sample and configured to house a collection bag (Col. 3, lines 1- 5). Neward discloses a stop mechanism (48) configured to reduce an inner diameter of the lumen. The collection bag is formed of silicone (Col. 3, lines 5 – 8).

Response to Arguments

4. Applicant's arguments with respect to the claims have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

5. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on


Art Unit: 3736


the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jonathan ML Foreman whose telephone number is (571)272-4724. The examiner can normally be reached on Monday - Friday 8:00 am - 4:30 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on (571)272-4726. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


JMLF


MAX F. HINDENBURG
PATENT EXAMINER
ELECTRONIC BUSINESS CENTER 3700